

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Dermatophyte Test Medium Slant

PRODUCT CODE NO: AS002-10

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Orange Red
pH	5.5 ± 0.2
Appearance	Sterile Dermatophyte Test Agar in disposable slants
Clarity	Clear to slightly opalescent
Sterility after 48 hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Cultural characteristics observed after incubation at 22 - 28°C for up to 6 days

Organism Description	Growth	Color of colony
Aspergillus brasiliensis ATCC 16404	[none-poor]	[-]
Candida albicans ATCC 10231	[good]	[-]
Microsporum audouinii ATCC 9079	[good]	[pink-red]
Pseudomonas aeruginosa ATCC 27853	[none-poor]	[-]
Trichophyton mentagrophytes ATCC 9533	[good]	[pink-red]
Trichophyton rubrum ATCC 28188	[good]	[pink-red]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com

Rev: January 2026