

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Gelatin Agar Slant**

**PRODUCT CODE NO: AS003-25**

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Yellow
pH	7.2 ± 0.2
Appearance	Sterile Kligler Iron Agar in disposable slants
Clarity	Clear to slightly opalescent
Sterility after 48 hrs	Sterile

### QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Cultural characteristics observed after incubation at 35-37 C for 24-48 hours.

Organism Description	Inoculum(CFU)	Growth	Recovery	Gelatin liquefaction
Vibrio cholerae ATCC 15748	[50-100]	[luxuriant]	[>=50%]	[positive reaction , clear zone around the colony within 24-48 hrs]
Vibrio parahaemolyticus ATCC 17802	[50-100]	[luxuriant]	[>=50%]	[positive reaction , clear zone around the colony within 24-48 hrs]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026