

CERTIFICATE OF ANALYSIS

PRODUCT : Tellurite blood agar base

PRODUCT CODE NO.: B085

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.2+/-0.2
COLOR (post autoclaving/heating)	a)Basal medium-Yellow.b)With addition of Haemoglobin-Reddish brown
CLARITY (post autoclaving/heating)	Opaque

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed along with added Haemoglobin solution (BF036), Vitamino Growth Supplement (BF037) and 1% Potassium Tellurite (BF010) after an incubation at 35-37°C for 48 hours (or more).

Organism Description	Inoculum (CFU)	Growth	Recovery	Colour of colony
Corynebacterium diphtheriae ATCC 11913	[50-100]	[Good - Luxuriant]	[>=50%]	[Grey-black]
Escherichia coli ATCC 25922	[>=10 ³]	[Inhibited]	[0%]	[-]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.