BIOMARK Laboratories – INDIA www.biomarklabs.com

CERTIFICATE OF ANALYSIS

PRODUCT: Kf streptococcal agar base

PRODUCT CODE NO.: B1077

QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES		
COLOR	Cream to light yellow		
APPEARANCE	Homogenous free flowing powder		

QUALITY CONTROL TESTS: REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH (post autoclaving/heating)	7.2+/-0.2
COLOR (post autoclaving/heating)	Cream to light yellow
CLARITY (post autoclaving/heating)	Clear to slightly opalescent

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed with added TTC solution 1% after 48-72 hrs at 35-37°C

Organism Description	INOCULUM(CFU)	GROWTH		COLOUR OF COLONEY
Enterococcus aerogenes (13048)	>=10³	inhibited	0%	-
Enterococcus faecalis (29212)	50-100	good-luxuriant	>=50%	Red-maroon
Escherichia coli (25922)	>=103	inhibited	0%	-

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.