CERTIFICATE OF ANALYSIS

QA/F/01 Rev no. 03

PRODUCT : Legionella agar base

PRODUCT CODE NO.: B1202

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Grey-Black
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	6.9+/-0.2
COLOR (post autoclaving/heating)	Black gel
CLARITY (post autoclaving/heating)	Opaque

QUALITY CONTROL TESTS : MICROBIOLOGICAL
CULTURAL RESPONSE:
Cultural characteristics observed with added Legionella Growth Supplement (FD016A),or Legionella Selective Supplement (FD017) and Legionella supplement (FD041A) after an incubation at 35-37°C for 48-72 hours.

Organism Description	Growth	Colour of colony
Legionella dumoffii ATCC33343	[Good-luxuriant]	[light blue to grey white]
Legionella pneumophila33153	[Good-luxuriant]	[light blue to grey white]