BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Maximum recovery diluent

PRODUCT CODE NO.: B1225

QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
COLOR	white to pale yellow	
APPEARANCE	Homogenous free flowing powder	

QUALITY CONTROL TESTS: REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.0+/-0.2
COLOR (post autoclaving/heating)	Light yellow
CLARITY (post autoclaving/heating)	Clear solution without any precipation

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed on Soyabean Casein Digest Agar (B039), after an incubation at 35-37°C for 18-24 hours of cultures suspended in Maximum Recovery Diluent for 30 minutes.

Organism Description	INOCULUM	RECOVERY AFTER 30 MIN.
Staphylococcus aureus ATCC 25923	[50 - 100]	[no change in numbers]
Escherichia coli ATCC 25922	[50 - 100]	[no change in numbers]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.