

BIOMARK LABORATORIES

CERTIFICATE OF ANALYSIS

PRODUCT : Mycoplasma urogenital broth base

PRODUCT CODE NO.: B1233

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Light yellow to pink
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	6.3 ± 0.2
COLOR (post autoclaving/heating)	Reddish pink
CLARITY (post autoclaving/heating)	clear solution

QUALITY CONTROL TESTS : MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37°C for 48 hours to one week with added Vitamino Growth Supplement (BF037), Urea Solution (BF124), Horse Serum (BF021) and Mycolasma Urogenital Selective Supplement (BF123).

Organism Description	GROWTH	Arginine	Urea
Mycoplasma hominis ATCC 14027	[Good -luxuriant]	[positive reaction, red colour]	[Negative reaction, no colour change]
Ureaplasma urealyticum ATCC 27618	[Good -luxuriant]	[negative reaction, no red colour]	[positive reaction, red orange colour]

This is to certify that this lot passes and it conforms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.