

## CERTIFICATE OF ANALYSIS

QA/F/01 Rev no. 03

**PRODUCT :** Oxacillin resistance screening agar base

**PRODUCT CODE NO.:** B1240

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Straw grey
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.2+/-0.2
COLOR (post autoclaving/heating)	Blue grey
CLARITY (post autoclaving/heating)	Slightly opalescent with slight precipitate

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed with added Oxacillin Resistance Selective Supplement , after an incubation at 35-37°C for 18-24 hours.

Organism Description	Growth
Staphylococcus aureus ATCC 29213	Inhibited (susceptible to oxacillin)
Staphylococcus aureus ATCC 38591	Fair to good (presence of colony or haze of growth should be read as resistance).