

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Brain heart infusion broth

**PRODUCT CODE NO.:** B1435

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Pale yellow to yellow
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.4+/-0.2
COLOR (post autoclaving/heating)	Light yellow to medium amber
CLARITY (post autoclaving/heating)	Clear

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours.

Organism Description	Inoculum (CFU)	Growth
Streptococcus pyogenes ATCC 19615	[50-100]	[Good - Luxuriant]
Escherichia coli (ATCC25922)	[50-100]	[Good - Luxuriant]
Streptococcus pneumoniae ATCC 6303	[50-100]	[Good - Luxuriant]
Staphylococcus aureus ATCC 25923	[50-100]	[Good - Luxuriant]

:

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid