

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Differential Reinforced Clostridial Agar

PRODUCT CODE NO.: B 1553

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.1+/-0.2
COLOR (post autoclaving/heating)	Light Pink
CLARITY (post autoclaving/heating)	clear to slightly opalescent gel

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed in an anaerobic atmosphere, after an incubation at 30-35°C for 1 week.

Organism Description	Inoculum (CFU)	Growth	Recovery	Colour of colony
Clostridium perfringens ATCC 13124	[50 - 100]	[Good - Luxuriant]	[>=50%]	[Black]
Clostridium sporogenes ATCC 11437	[50 - 100]	[Good - Luxuriant]	[>=50%]	[Black]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: December 2023