

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Enterococcus confirmatory agar

**PRODUCT CODE NO.:** B186

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Yellow
APPEARANCE	Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	8.0+/-0.2
COLOR (post autoclaving/heating)	Light blue
CLARITY (post autoclaving/heating)	Clear to slightly opalescent

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 25° C for 48 hours.

Organism Description	Inoculum (CFU)	Growth
Escherichia coli ATCC 25922	$\geq 10^3$	inhibited
Enterococcus faecalis ATCC 29212	50-100	good-luxuriant

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.