

CERTIFICATE OF ANALYSIS

PRODUCT : Lethen broth,aoac

PRODUCT CODE NO.: B224

QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS: REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH (post autoclaving/heating)	7.0+/-0.2
COLOR (post autoclaving/heating)	Light yellow
CLARITY (post autoclaving/heating)	Clear solution

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours.

Organism Description	Inoculum (CFU)(Standard)	Growth (Standard)
Escherichia coli ATCC 25922	50-100	Good-luxuriant
Staphylococcus aureus ATCC 6538	50-100	Good-luxuriant
Escherichia coli ATCC 8739	50-100	Good-luxuriant
Staphylococcus aureus ATCC 25923	50-100	Good-luxuriant

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.