#### BIOMARK Laboratories – INDIA www.biomarklabs.com

# **CERTIFICATE OF ANALYSIS**

**PRODUCT :** Malt extract agar base (w/mycological peptone)

PRODUCT CODE NO.: B249

## **QUALITY CONTROL TESTS : DEHYDRATED MEDIUM**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
COLOR	Light to medium yellow	
APPEARANCE Homogenous free flowing powder		

### **QUALITY CONTROL TESTS : REHYDRATED MEDIUM**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
pH(post autoclaving/heating)	5.4+/-0.2	
COLOR (post autoclaving/heating)	Light Brown	
CLARITY (post autoclaving/heating)	Clear to slightly opalescent	

# QUALITY CONTROL TESTS : MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 25-30°C for 48-72 hrs.

Organism Description	Inoculum(CFU)	Growth	Recovery
Aspergillus brasilliensis ATCC 16404	50-100	luxuriant	-
Candida albicans ATCC 10231	50-100	luxuriant	>=70%
Saccharomyces cerevisiae ATCC9763	50-100	luxuriant	>=70%

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.