

CERTIFICATE OF ANALYSIS

PRODUCT : Phosphate buffer,apha,ph7.2

PRODUCT CODE NO.: B297

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Offwhite to white
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.2+/-0.2
COLOR (post autoclaving/heating)	Colourless
CLARITY (post autoclaving/heating)	clear solution

QUALITY CONTROL TESTS : MICROBIOLOGICAL  
CULTURAL RESPONSE:

Cultural characteristics observed after incubation at (Drop down) for (Drop down) hours.

Organism Description

Not applicable

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.