

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT :** Acetamide agar (twin pack)

**PRODUCT CODE NO.:** B910

### QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Part A- Colourless.Part B-pink
APPEARANCE	Part A-Crystals.Part B-Homogenous free flowing powder

### QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	6.9+/-0.2
COLOR (post autoclaving/heating)	Orange
CLARITY (post autoclaving/heating)	Clear to slightly opalescent

### QUALITY CONTROL TESTS : MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37° C for 4-7 days.

Organism Description	Inoculum (CFU)	Growth	Deamination
Stenotrophomonas maltophilia ATCC 13637	[50-100]	[good-luxuriant]	[negative reaction,no purplish red colour within 7 days]
Pseudomonas aeruginosa ATCC 27853	[50-100]	[good-luxuriant]	[positive reaction, purplish red colour within 7 days]

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.