

CERTIFICATE OF ANALYSIS

QA/F/01 Rev no. 03

PRODUCT : Acetate agar
PRODUCT CODE NO.: B912

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Beige
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	5.4+/-0.2
COLOR (post autoclaving/heating)	Cream to yellow
CLARITY (post autoclaving/heating)	Clear to slightly opalescent

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 25-30°C for 18-48 hours.

Organism Description	Growth
Enterococcus faecalis ATCC 29212	[none-poor]
Leuconostoc mesenteroides ATCC 12291	[good-luxuriant]
Pediococcus acidilactici ATCC33314	[good-luxuriant]