

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Blood free campylobacter selectivity agar base

PRODUCT CODE NO.: B945

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Grey
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	7.4+/-0.2
COLOR (post autoclaving/heating)	Black
CLARITY (post autoclaving/heating)	Opaque

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed with added Campylobacter Supplement V(BFCSA)(BF018) & CAT selective supplement (BF108), after an incubation 42°C for 24-48 hours.

Organism Description	Inoculum (CFU)	Growth	Colour of colony
Campylobacter coli (ATCC 33559)	[50-100]	[Good-Luxuriant]	[Creamy-grey]
Campylobacter jejuni (ATCC 29428)	[50-100]	[Good-Luxuriant]	[grey]
Campylobacter laridis (ATCC 35222)	[50-100]	[Good-Luxuriant]	[varying type]
Escherichia coli (ATCC 25922)	[>=10 ³]	[Inhibited]	[-]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: December 2023