

**CERTIFICATE OF ANALYSIS**

**PRODUCT: ZN ACID FAST STAINS KIT (CONTAINS BA007, BA008, BA009)**  
**PRODUCT CODE NO: BA006**

**QUALITY CONTROL: PHYSICAL**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
<b>COLOUR</b>	BA007 -Colourless BA008- Dark pinkish red solution with sheen BA009- Dark blue
<b>APPEARANCE</b>	Liquid

**QUALITY CONTROL: MICROSCOPY (STAINING)**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
<b>MICROSCOPY</b>	BA007- Decolouriser for non acid fast bacteria. BA008- An acid fast organism stains bright red BA009- Metachromatic granules are stained deep blue and cytoplasm pale blue In acid fast staining nonacid fast organisms and cellular material are stained blue.

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.