

**CERTIFICATE OF ANALYSIS**

**PRODUCT: METHYLENE BLUE (LOFFLER'S)**

**PRODUCT CODE NO: BA009**

**QUALITY CONTROL: PHYSICAL**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Dark Blue
APPEARANCE	Liquid
CLARITY	Clear

**QUALITY CONTROL: MICROSCOPY (STAINING)**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
MICROSCOPY	Metachromatic granules are stained deep blue and cytoplasm pale blue.
MICROSCOPY	In acid fast staining nonacid fast organisms and cellular material are stained blue.

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.