

CERTIFICATE OF ANALYSIS

PRODUCT: BROMOPHENOL BLUE INDICATOR SOLUTION

PRODUCT CODE NO: BA049

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Bluish-violet coloured solution
APPEARANCE	Liquid
CLARITY	Clear

QUALITY CONTROL: MICROSCOPY (STAINING)

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
REACTION	Colour change: Yellow to blue At pH 3.0 the indicator turns yellow and at pH 4.6 the indicator is blue

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.