

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : BENEDICTS QUALITATIVE REAGENT
PRODUCT CODE BA060

QUALITY PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD
Colour	Blue
Appearance	Liquid
Clarity	Clear
Colour	Blue coloured solution.
Result	-
Blue colour –	Sugar absent
Green and slight yellow	Sugar present
Yellow and orange-	Sugar present
Brick red and orange red-	Sugar present

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Web Site:-** www.biomarklabs.com

Rev: January 2025