

CERTIFICATE OF ANALYSIS

PRODUCT : EHR LIC ALDEHYDE REAGENT
PRODUCT CODE BA064

QUALITY PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD
COLOUR	Light yellow
APPEARANCE	Liquid
CLARITY	Clear solution with characteristic odour
Test Reaction Observed By Addition Of 1ml Of Reagent To 10ml Of Urine.	
1.Cherry red colour	Increased urobilinogen
2.Absence of colour	Deareased or normal amount of urobilinogen

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.