

CERTIFICATE OF ANALYSIS

PRODUCT : KOVACS INDOLE REAGENT

PRODUCT CODE BA068

QUALITY CONTROL:**PHYSICAL**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Color	Greenish yellow
Appearance	Liquid
Clarity	Clear

QUALITY CONTROL: MICROBIOLOGICAL

Cultural characteristics observed after 18 - 24 hrs at 35-37°C After addition in Tryptone water.

ORGANISM DESCRIPTION	INDOLE REACTION
Escherichia coli (25922)	Positive, red ring at the interface of the medium
Enterobacter aerogenes (13048)	Negative, no red ring

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.