

CERTIFICATE OF ANALYSIS

QA/F/63/ Rev No. 00

<b>PRODUCT :</b>	Kovacs indole reagent
<b>PRODUCT CODE NO.:</b>	BA068

**QUALITY CONTROL:** Physical

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Color	Light yellow
Appearance	Liquid
Clarity	Clear

**QUALITY CONTROL : MICROBIOLOGICAL**

Cultural characteristics observed after 18 - 24 hrs at 35-37°C After addition in Tryptone water.

Organism Description	INDOLE REACTION
Escherichia coli.(25922)	Positive
Enterobacter aerogenes(13048)	Negative