

CERTIFICATE OF ANALYSIS

PRODUCT: SULPHOSALICYLIC ACID 3%

PRODUCT CODE NO: BA079

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Colourless May have slight pink tinge
APPEARANCE	Liquid
CLARITY	Clear

QUALITY CONTROL: MICROSCOPY (STAINING) - NA

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
RESULT	Protein from urine can be estimated by the turbidimetric method using sulphosalicylic acid reagent.
	7.5 mg/100ml Negative
	up to 20mg/100ml Trace
	up to 30mg/100ml +
	up to 50mg/100ml ++
	up to 75mg/100ml +++

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.