

CERTIFICATE OF ANALYSIS

QA/F/63/ Rev No. 00

PRODUCT : Barrit reagent a**PRODUCT CODE NO.:** BA084**QUALITY CONTROL:** Physical

| PARAMETERS | SPECIFICATIONS/STANDARD VALUES |
|------------|--------------------------------|
| Colour | Colourless to reddish brown |
| Appearance | Liquid |
| Clarity | Clear |

QUALITY CONTROL: MICROSCOPY (STAINING)

| PARAMETERS | SPECIFICATIONS/STANDARD VALUES |
|-----------------------------------|--|
| Cultural response | Biochemical identification was carried out by adding 0.6 ml of Barritt Reagent (Part A) and 0.2 ml |
| Organism | Growth & VP Test |
| Enterobacter aerogenes ATCC 13048 | Luxuriant Positive (Cherry red colour formation) |
| Escherichia coli ATCC 25922 | Luxuriant Negative (No red colour formation) |
| Klebsiella pneumoniae ATCC 13883 | Luxuriant Positive (Cherry red colour formation) |