

## CERTIFICATE OF ANALYSIS

QA/F/63/ Rev No. 00

**PRODUCT :** Barrit reagent b**PRODUCT CODE NO.:** BA085**QUALITY CONTROL:** Physical

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Colourless
Appearance	Liquid
Clarity	Clear

**QUALITY CONTROL: MICROSCOPY (STAINING)**

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Cultural response	Biochemical identification was carried out by adding 0.2 ml of Barritt Reagent (Part A) and 0.2 ml of Barritt Reagent (Part B)
Organism :	Growth & VP Test
Enterobacter aerogenes ATCC 13048	Luxuriant Positive (Cherry red colour formation)
Escherichia coli ATCC 25922	Luxuriant Negative (No red colour formation)
Klebsiella pneumoniae ATCC 13883	Luxuriant Positive (Cherry red colour formation)