

CERTIFICATE OF ANALYSIS

PRODUCT: GORDEN-MCLEOD REAGENT (OXIDASE REAGENT)

PRODUCT CODE NO: BA101

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Black
APPEARANCE	Liquid
CLARITY	Clear to slightly hazy liquid

QUALITY CONTROL: MICROSCOPY (STAINING)

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Organism	Oxidase test
Pseudomonas aeruginosa (27853)	Positive (development of purple-blue colour)
Staphylococcus aureus (25923)	Negative (No colour change)
Neisseria gonorrhoeae (19424)	Positive (development of purple-blue colour)

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.