

CERTIFICATE OF ANALYSIS

PRODUCT: O'MEARA REAGENT

PRODUCT CODE NO: BA104

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Colourless
APPEARANCE	Liquid
CLARITY	Clear

QUALITY CONTROL: MICROSCOPY (STAINING)

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
Organism	GROWTH	VP TEST
Escherichia coli (25922)	Luxuriant	Negative(No red colour formation)
Enterobacter aerogenes (13048)	Luxuriant	Positive(Red colour formation)
Klebsiella pneumoniae (13883)	Luxuriant	Positive(Red colour formation)

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.