

# BIOMARK LABORATORIES

## CERTIFICATE OF ANALYSIS

QA/F/62/Rev No.: 00

<b>PRODUCT:</b> Aztreonam 30 mcg <b>PRODUCT CODE NO.:</b> BD007
--------------------------------------------------------------------

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Filter paper discs of 6mm diameter with <b>AT</b> printed on discs.

### QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours.

Antimicrobial Disc with concentration	Test Organisms (ATCC)	Standard diameter of zone of inhibition in mm		
		Resistant	Intermediate	Sensitive
Aztreonam 30 mcg	E.coli ATCC 25922	≤ 17	18 -20	≥ 21
	Ps.aeruginosa ATCC 27853	≤ 15	16 - 21	≥ 22