

CERTIFICATE OF ANALYSIS

PRODUCT: Cefaclor 30 mcg

PRODUCT CODE NO.: BD010

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
APPEARANCE	Filter paper discs of 6mm diameter with CFC printed on discs	

QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours.

Antimicrobial Disc with concentration	Test Organisms (ATCC)	Standard diameter of zone of inhibition in mm		
		Resistant	Intermediate	Sensitive
Cefaclor 30 mcg	E.coli (25922)	≤ 14	15 -17	≥ 18
	S.aureus(25923)	≤ 14	15 -17	≥ 18