

# BIOMARK LABORATORIES

## CERTIFICATE OF ANALYSIS

QA/F/62/Rev No.: 00

**PRODUCT: Clindamycin 2 mcg**

**PRODUCT CODE NO.: BD028**

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Filter paper discs of 6mm diameter with CD printed on discs.

### QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours.

Antimicrobial Disc with concentration	Test Organisms (ATCC)	Standard diameter of zone of inhibition in mm		
		Resistant	Intermediate	Sensitive
Clinadamycin 2 mcg	S.aureus ATCC 25923	≤ 14	15 -20	≥ 21