

CERTIFICATE OF ANALYSIS

<p>PRODUCT: Levofloxacin 5mcg</p> <p>PRODUCT CODE NO.: BD041</p>
--

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
APPEARANCE	Filter paper discs of 6mm diameter with LV printed on discs	

QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours.

Antimicrobial Disc with concentration	Test Organisms (ATCC)	Standard diameter of zone of inhibition in mm		
		Resistant	Intermediate	Sensitive
Levofloxacin (5 mcg)	E.coli (25922)	≤ 13	14-16	≥17
	S.aureus (25923)	≤ 13	14-16	≥17
	P.aeruginosa (27853)	≤ 13	14-16	≥17