

BIOMARK LABORATORIES - PUNE

CERTIFICATE OF ANALYSIS

PRODUCT:MOXALACTAM (MX 30 mcg)

PRODUCT CODE NO.: BD 047

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	White discs of 6 mm diameter with abbreviated name MX

QUALITY CONTROL: MICROBIOLOGICAL

Measure the zone diameters in mm observed in the plates and compare them with the CLSI zone chart, as applicable. Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours as per CLSI / NON CLSI standards.

Micro-organism with ATCC number Moxalactam (MX 30 Mcg)	Diameter of zone of inhibition
	Quality control limit in mm
EC (ATCC 25922)	28-35
SA (ATCC 25923)	18-24
PS (ATCC 27853)	17-25

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Web Site:-** www.biomarklabs.com

Rev: January 2025