

CERTIFICATE OF ANALYSIS

<p>PRODUCT: Amoxyclav 30mcg</p> <p>PRODUCT CODE NO.: BD078</p>
--

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
APPEARANCE	Filter paper discs of 6mm diameter with AU printed on discs	

QUALITY CONTROL: MICROBIOLOGICAL

Average diameter of zone of inhibition in mm observed on Mueller Hinton Agar Plates (B263) at 35-37°C for 18-24 hours.

Antimicrobial Disc with concentration	Test Organisms (ATCC)	Standard diameter of zone of inhibition in mm		
		Resistant	Intermediate	Sensitive
Amoxyclav (30 mcg)	E.coli (25922)	≤ 13	14-17	≥18
	S.aureus (25923)	≤ 19	-	≥20