

BIOMARK LABORATORIES

CERTIFICATE OF ANALYSIS

QA/F/62/Rev No.: 00

PRODUCT:Bacitracin paper discs

PRODUCT CODE NO.: BDD022

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Filter paper discs of 6mmdiameter

QUALITY CONTROL: MICROBIOLOGICAL

On Tryptose Blood agar plates at 35-37°C for 18-24 hours

Organism(ATCC)	Reaction Diameter of zone of Inhibition
Streptococcus pyogenes (19615)	15mm