

CERTIFICATE OF ANALYSIS

PRODUCT: Bile Esculin

PRODUCT CODE NO.: BDD023

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/ STANDARD VALUES	
APPEARANCE	Filter paper discs of 6mm diameter	

QUALITY CONTROL: MICROBIOLOGICAL

Cultural characteristics on the seeded Bile Esculin Agar Base (B424) plate or another media after 18-24 hours at 35-37°C.

Organisms (ATCC)	Esculin Hydrolysis
Streptococcus faecalis (29212)	+
Streptococcus pyrogenes (19615)	-
Listeria monocytogenes (19118)	+