

# BIOMARK LABORATORIES

## CERTIFICATE OF ANALYSIS

QA/F/62/Rev No.: 00

**PRODUCT: Oxidase discs**

**PRODUCT CODE NO.: BDD027**

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Filter paper discs of 6mmdiameter

### QUALITY CONTROL: MICROBIOLOGICAL

Typical oxidase reaction given by 18-48-hourculture within 5-10 seconds at 25-30°C.

Organism(ATCC)	Reaction
Pseudomonas aeruginosa (27853)	Positive: Deep purplish blue colourations of disc
Staphylococcus aureus(25923)	Negative: No colour change
Escherichia coli (259222)	Negative: No colour change