

## CERTIFICATE OF ANALYSIS

**PRODUCT: Sterile Disc 6mm**

**PRODUCT CODE NO.: BDD028**

### QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES	
APPEARANCE	Sterile Filter paper discs of 6mm diameter	

### QUALITY CONTROL: MICROBIOLOGICAL

Sterility of the discs checked for 14 days in Alternative Thioglycollate medium (B011) at 30-35°C and Soyabean casein digest medium (B040) at 20-25°C.

Sample	Result in Alternative Thioglycollate medium (B011)	Result in Soyabean casein digest medium (B040)
Test	Clear, No growth	Clear, No growth
Negative Control	Clear, No growth	Clear, No growth
Positive Control	Luxuriant growth	Luxuriant growth