

BIOMARK LABORATORIES

CERTIFICATE OF ANALYSIS

PRODUCT: OXIDASE STRIPS

PRODUCT CODE NO.: BDD031

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
APPEARANCE	Paper strips of size 60 x 5 mm

QUALITY CONTROL: MICROBIOLOGICAL

The reaction is observed within 5-10 seconds at 25-30°C. A delayed positive reaction appears in 10-60 seconds whereas a change later than 60 seconds or no change is considered negative reaction.

ORGANISM DESCRIPTION (ATCC)	Reaction	Colour
Escherichia coli (259222)	Negative	No colour change
Neisseria gonorrhoeae (19424)	Positive	Deep purple blue
Staphylococcus aureus(25923)	Negative	No colour change
Pseudomonas aeruginosa (27853)	Positive	Deep purple blue

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com.

Rev: January 2025