BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT : Fraser selective supplement

PRODUCT CODE NO: BF117I

QUALITY CONTROL:

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Light brown
APPEARANCE	Powder
CLARITY	NA

QUALITY CONTROL : MICROBIOLOGICAL
CULTURAL RESPONSE:

Cultural characteristics observed after (24-48) hours at (35-37)oC when added to FRASER BROTH BASE(B963I) along with Fraser Supplement (BF002)

Orgm Desc	INOCULUM(CFU)	GROWTH	ESCULIN HYDROLYSIS
Escherichia coli (25922)	>=10^3	Inhibited	-
Enterococcus faecalis (29212)	>=10^3	Inhibited	-
Listeria monocytogenes (19111)	50-100	good-luxuraiant	positive reaction,blackening of medium.
Listeria monocytogenes (19112)	50-100	good-luxuraiant	positive reaction, blackening of medium.
Listeria monocytogenes (19117)	50-100	good-luxuraiant	positive reaction,blackening of medium.
Listeria monocytogenes (19118)	50-100	good-luxuraiant	positive reaction,blackening of medium.
Staphylococcus aureus (25923)	>=10^3	Inhibited	-

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid