

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Sabouraud –glucose agar with chloramphenicol

PRODUCT CODE NO.: BH709

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	5.6+/-0.2
COLOR (post autoclaving/heating)	Light amber
CLARITY (post autoclaving/heating)	Clear to slightly opalescent

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 20-25°C for 48-72 hours (Incubate for 7 days for Trychophyton species).

Organism Description	Inoculum(CFU)	Growth	Recovery
Aspergillus brasiliensis ATCC 16404	[50-100]	[good-luxuriant]	[-]
Candida albicans ATCC 10231	[50-100]	[good-luxuriant]	[>=50%]
Escherichia coli ATCC 25922	[>=10 ³]	[inhibited]	[0%]
Lactobacillus casei ATCC 334	[>=10 ³]	[inhibited]	[0%]
Saccharomyces cerevisiae ATCC 9763	[50-100]	[good-luxuriant]	[>=50%]
Trichophyton rubrum ATCC ATCC 28191	[50-100]	[good-luxuriant]	[-]
Escherichia coli NCTC 9002	[>=10 ³]	[inhibited]	[0%]
Escherichia coli ATCC 8739	[>=10 ³]	[inhibited]	[0%]

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **WebSite:-** www.biomarklabs.com

Rev: January 2026