

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: DILUTING FLUID, A (PEP 0.1%)
PRODUCT CODE NO: BR003

QUALITY CONTROL: PHYSICAL

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
Colour	Light yellow coloured medium
pH	7.1 ± 0.2
Appearance	Sterile clear Diluting Fluid A in bottle.
Clarity	Clear
Sterility after 48 hrs	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE

Cultural characteristics observed after incubation at 35-37°C for 18-24 hours

Organism Description	Inoculum(CFU)	Growth	Recovery
Escherichia coli 25922	[50-100]	[luxuriant]	[>=50%]
Candida albicans 10231	[50-100]	[luxuriant]	[>=50%]
Staphylococcus aureus 25923	[50-100]	[luxuriant]	[>=50%]
Staphylococcus aureus 6538	[50-100]	[luxuriant]	[>=50%]
Escherichia coli 8739	[50-100]	[luxuriant]	[>=50%]
Pseudomonas aeruginosa 9027	[50-100]	[luxuriant]	[>=50%]
Aspergillus niger 16404	[50-100]	[luxuriant]	[>=50%]
Bacillus subtilis 6633	[50-100]	[luxuriant]	[>=50%]
Clostridium sporogenes 19404	[50-100]	[luxuriant]	[>=50%]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, Website:- www.biomarklabs.com

Rev: January 2026