

CERTIFICATE OF ANALYSIS

QA/F/01 Rev no. 03

PRODUCT : Antibiotic assay medium no. 32

PRODUCT CODE NO.: BS030

QUALITY CONTROL TESTS : DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOR	Cream to Light yellow
APPEARANCE	Homogenous free flowing powder

QUALITY CONTROL TESTS : REHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
pH(post autoclaving/heating)	6.6+/-0.2
COLOR (post autoclaving/heating)	Light to medium yellow
CLARITY (post autoclaving/heating)	Clear to slightly opalescent

QUALITY CONTROL TESTS : MICROBIOLOGICAL

CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37° C for 5 days.

Organism Description	Inoculum (CFU)	Growth	Recovery	Antibiotics assayed
Bacillus subtilis ATCC 6633	[50-100]	[luxuriant]	[>=70%]	[Monensin]