

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: BILE BROTH**

**PRODUCT CODE NO: LM002-25**

### QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

| PARAMETERS                      | SPECIFICATIONS/STANDARD VALUES |
|---------------------------------|--------------------------------|
| COLOUR                          | Light yellow                   |
| APPEARANCE                      | clear Bile Broth in vials      |
| STERILITY TEST (AFTER 48 HOURS) | Sterile                        |

### QUALITY CONTROL TESTS: MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after incubation at 35-37°C for 18-48 hours

| Organism Description         | Inoculum(CFU)        | Growth       |
|------------------------------|----------------------|--------------|
| Escherichia coli ATCC 25922  | [50-100]             | [Luxuriant ] |
| Enterobacter aerogenes 13048 | [50-100]             | [Luxuriant ] |
| Salmonella typhimurium 14028 | [50-100]             | [Luxuriant ] |
| Staphylococcus aureus 25923  | [>=10 <sup>4</sup> ] | [Inhibited ] |

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026