

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Koser's Citrate Medium.**

**PRODUCT CODE NO: LM007-50**

### QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Colorless, clear medium
APPEARANCE	Sterile Koser's Citrate Medium in glass vial
STERILITY TEST (AFTER 48 HOURS)	Sterile

### QUALITY CONTROL TESTS: MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after incubation at 35-37°C for 18-24 hours

Organism Description	Inoculum(CFU)	Growth	CITRATE UTILISATION
Escherichia coli 25922	[50-100]	[None-poor]	[Negative]
Enterobacter aerogenes 13048	[50-100]	[Luxuriant ]	[Positive]
Enterobacter cloacae 23355	[50-100]	[Luxuriant ]	[Positive]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026