

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: MR-VP Broth (Glucose Phosphate Broth)**

**PRODUCT CODE NO: LM013-25**

### QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Light yellow colored clear solution.
APPEARANCE	Sterile clear MR-VP Broth in glass bottle.
STERILITY TEST (AFTER 48 HOURS)	Sterile

### QUALITY CONTROL TESTS: MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after an incubation at 35-37°C for 24-48 hours.

Organism Description	Inoculum(CFU)	Growth	VP TEST	MR TEST
Escherichia coli 25922	[50-100]	[Luxuriant ]	Negative reaction, yellow or no color change	Positive reaction, bright red color
Enterobacter aerogenes 13048	[50-100]	[Luxuriant ]	Positive reaction, eosin pink/red color	Negative reaction, yellow or no color change
Klebsiella pneumoniae 23357	[50-100]	[Luxuriant ]	Positive reaction, eosin pink/red color	Negative reaction, yellow or no color change

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026