

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Selenite Broth

PRODUCT CODE NO: LM019-25

QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Light yellow coloured medium glass bottle.
APPEARANCE	Sterile Clear Selenite Broth in glass bottle.
STERILITY TEST (AFTER 48 HOURS)	Sterile

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after incubation at 35-37°C for 18-24 hours when recovered on MacConkey Agar (B236).

Organism Description	Inoculum(CFU)	Recovery
S. SEROTYPE TYPHIMURIUM 14028	[50-100]	Good-luxuriant
Escherichia coli 25922	[$\geq 10^3$]	No increase in number
Salmonella Typhi 6539	[50-100]	Good-luxuriant
Salmonella Choleraesuis 12011	[50-100]	Good-luxuriant

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.
Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com

Rev: January 2026