

# BIOMARK LABORATORIES, INDIA

## CERTIFICATE OF ANALYSIS

**PRODUCT: Brucella Agar Plate with Hemin & Vitamin K1**

**PRODUCT CODE NO: PP003-20**

### QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

| PARAMETERS             | SPECIFICATIONS/STANDARD VALUES   |
|------------------------|--|
| COLOUR                 | Red coloured medium  |
| APPEARANCE             | Brain Brucella agar w/hemin and vitamin K1 in 90X15mm disposable plates. |
| CLARITY                | Slightly opalescent  |
| STERILITY AFTER 48 HRS | Sterile  |
| PH                     | 7.0 ± 0.2  |

### QUALITY CONTROL TESTS: MICROBIOLOGICAL

#### CULTURAL RESPONSE:

Cultural characteristics observed after incubation at 35-37°C for 18-24 hours.

| Organism Description          | Inoculum(CFU) | Growth      | Recovery |
|-------------------------------|---------------|-------------|----------|
| Bacteroides fragilis 2528     | [50-100]      | [Luxuriant] | [>=50%]  |
| Clostridium perfringens 13124 | [50-100]      | [Luxuriant] | [>=50%]  |

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

**Address:-** S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

**Email ID:-** biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** [www.biomarklabs.com](http://www.biomarklabs.com)

Rev: January 2026