

BIOMARK LABORATORIES, INDIA

CERTIFICATE OF ANALYSIS

PRODUCT: Chocolate Agar Plate

PRODUCT CODE NO: PP005-50

QUALITY CONTROL TESTS: DEHYDRATED MEDIUM

PARAMETERS	SPECIFICATIONS/STANDARD VALUES
COLOUR	Chocolate brown coloured medium.
APPEARANCE	Chocolate Agar in 90X15mm disposable plates.
CLARITY	Slightly opalescent
STERILITY AFTER 48 HRS	Sterile
PH	7.3 ± 0.2

QUALITY CONTROL TESTS: MICROBIOLOGICAL CULTURAL RESPONSE:

Cultural characteristics observed after incubation at 35-37°C for 24-48 hours..

Organism Description	Inoculum(CFU)	Growth	Recovery
Neisseria gonorrhoeae 19424	[50-100]	[Luxuriant]	[>=50%]
Neisseria meningitidis 13090	[50-100]	[Luxuriant]	[>=70%]
Streptococcus pneumoniae 6303	[50-100]	[Luxuriant]	[>=50%]
Streptococcus pyogenes 19615	[50-100]	[Luxuriant]	[>=50%]
Haemophilus influenzae 19418	[50-100]	[Luxuriant]	[>=50%]

This is to certify that this lot passes and it confirms to the above-mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current Biomark literature. The results reported were obtained at the time of release.

This document has been produced electronically and is valid.

Address:- S.NO.135/1, DALVI WADI, NEAR NANDED PHATA, SINHAGAD ROAD, DHAIRI, PUNE-411041, INDIA.

Email ID:- biomarklabs@gmail.com / biomarkmktg@gmail.com, **Website:-** www.biomarklabs.com

Rev: January 2026